



Türk Akreditasyon Kurumu

ACCREDITATION CERTIFICATE

Operating as a Calibration Laboratory,

ENELSAN ENDÜSTRİYEL ELEKTRONİK SANAYİ ANONİM ŞİRKETİ

Merkez Adres: ÇERKEŞLİOSB MAH. IMES 4 BLV. ENELSAN NO:11 DİLOVASI Kocaeli / Türkiye

As a result of the audit conducted by TÜRKAK, it has been accredited in the scopes included in the Annex according to the TS EN ISO / IEC 17025: 2017 standard.

Accreditation No. : AB-0106-K

Accreditation Date : 13.01.2014

Revision Date / No : 01.09.2022 / 05



This Certificate is valid until 09.05.2026 if the above-mentioned organization continues to comply with the TS EN ISO/IEC 17025:2017 Standard, relevant Regulations and Communiqués.

Gülden Banu Müderrisoğlu
Genel Sekreter



The Turkish Accreditation Agency (TÜRKAK) has signed a multilateral agreement (MLA/MRA) with the European Accreditation Association (EA) and the International Laboratory Accreditation Association (ILAC) in the field of ISO/IEC 17025.

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. You can use the QR code to verify the e-signed document.

 TÜRKAK  Kalibrasyon TS EN ISO/IEC 17025 AB-0106-K	ENELSAN ENDÜSTRİYEL ELEKTRONİK SANAYİ ANONİM ŞİRKETİ Accreditation No: AB-0106-K Revision Date / No : 01.09.2022
Calibration Laboratory	
Address : ÇERKEŞLİ OSB MAH. İMES 4 BLV. ENELSAN NO:11 DİLOVAŞI Kocaeli / Türkiye	Tel: : +90 262 754 6313 Fax : - E-Mail : mustafa@enelsan.com Web address: :

Calibration and Measurement Capability (CMC)

Fluid Sizes

Measured Magnitude / Calibrated Devices	Measuring Range	Measurement Conditions	Extended Measurement Uncertainty (k=2)	Remarks / Calibration Method
Volumetric Liquid Flows Volumetric Water Flows Water meter	1 m ³ /h < Q < 1000 m ³ /h	(15-30) °C Su	%0,23	Q: Measured flow value <ul style="list-style-type: none"> At the Customer's Site In the Laboratory In late or mobile facilities ~calibration is done. Comparison Method with Reference Flowmeter System Applied standard: TS EN ISO 4064-2 Article 7.4
Mass Liquid Flows Mass Water Debs Water meter	1 m ³ /h < Q < 15 m ³ /h	(15-30) °C Su	%0,17	Q: Measured flow value in the lab <ul style="list-style-type: none"> cal bration is done. Comparison Method with Mass Flow Measurement System Applied standard: TS EN ISO 4064-2 Article 7.4
Mass Liquid Flows Mass Water Debs Water meter	15 m ³ /h < Q < 700 m ³ /h	(15-30) °C Su	%0,17	Q: Measured flow value in the lab <ul style="list-style-type: none"> cal bration is done. Comparison Method with Mass Flow Measurement System Applied standard: TS EN ISO 4064-2 Article 7.4
Mass Liquid Flows Mass Water Debs Water meter	700 m ³ /h < Q < 1000 m ³ /h	(15-30) °C Su	%0,20	Q: Measured flow value in the lab <ul style="list-style-type: none"> cal bration is done. Comparison Method with Mass Flow Measurement System Applied standard: TS EN ISO 4064-2 Article 7.4

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. You can use the QR code to verify the e-signed document.

